

Declaration of Conformity in accordance with ISO/IEC 17050-1

Document number	24-002 ISO
Issuer's Name	ATTO Corporation
Issuer's Address	2-21-6 Taito, Taito-ku, Tokyo 110-0016
Products of the Declaration	Chemiluminescence imaging devices (see appendix)

The subject of the above declaration complies with the requirements of the following documents:

Document Number	Name / Title
ISO 24421:2023	Biotechnology - Minimum requirements for optical signal measurements in photometric methods for biological samples
Edition/Publication date	Clause 5.2 Verification of optical signal measurement instruments
1st edition 2023-07	

Additional Information

Validation of declared products	Verify the relative sensitivity from images taken of an optical reference (reference LED light source) with a specified total radiant flux (nW).
---------------------------------	--

Contact Us	2-21-6 Taito, Taito-ku, Tokyo 110-0016, JAPAN ATTO Corporation Management Department Quality Assurance Office TEL: +81-3-5818-7560
Publication date	June 14, 2024
Place of issue	2-21-6 Taito, Taito-ku, Tokyo 110-0016, JAPAN ATTO Corporation Management Department Quality Assurance Office
Representative/Agent Position, Name	Quality Assurance Department Manager Agent: Naohito Omata Agent Signature <u>Naoto Omata</u>

*This document is a self-declaration of conformity based on ISO/IEC 17050-1 (JIS Q 17050-1).



ATTO Corporation, Management Department Quality Assurance Office

The products covered by this declaration are as follows:

Reference LED light source #1: ATTO WSL-1200/1200C KohshiFundam Reference

LED light source

name	Model	
LuminoGraph I CMOS	WSE-6170	Compliance verified by using reference LED light source #1
LuminoGraph II EM	WSE-6270	Compliance verified by using reference LED light source #1
LuminoGraph III Lite	WSE-6370	Compliance verified by using reference LED light source #1